

## King's Research Portal

DOI:

[10.1016/j.nima.2019.162365](https://doi.org/10.1016/j.nima.2019.162365)

Document Version

Peer reviewed version

[Link to publication record in King's Research Portal](#)

*Citation for published version (APA):*

Suhling, K., Hirvonen, L. M., Becker, W., Smietana, S., Netz, H., Milnes, J., Conneely, T., Le Marois, A. M., Jagutzki, O., Fred, F., Petrášek, Z., & Beeby, A. (2019). Wide-field Time-correlated Single Photon Counting-based Fluorescence Lifetime Imaging Microscopy. In *Nuclear Instruments and Methods A* (Vol. 942). [162365] (Nuclear Instruments and Methods in Physics Research, Section A: Accelerators, Spectrometers, Detectors and Associated Equipment). Elsevier. <https://doi.org/10.1016/j.nima.2019.162365>

### Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

### General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

### Take down policy

If you believe that this document breaches copyright please contact [librarypure@kcl.ac.uk](mailto:librarypure@kcl.ac.uk) providing details, and we will remove access to the work immediately and investigate your claim.

Manuscript Number: NIMA\_PROCEEDINGS-D-19-00021R1

Title: Wide-field Time-correlated Single Photon Counting-based  
Fluorescence Lifetime Imaging Microscopy

Article Type: VSI: ULITIMA 2018

Keywords: time-correlated single photon counting  
microchannel plate  
fluorescence lifetime imaging

Corresponding Author: Professor klaus suhling, PhD

Corresponding Author's Institution: king's college london

First Author: klaus suhling, PhD

Order of Authors: klaus suhling, PhD

Abstract: Wide-field time-correlated single photon counting detection techniques, where the position and the arrival time of the photons are recorded simultaneously using a camera, have made some advances recently. The technology and instrumentation used for this approach is employed in areas such as nuclear science, mass spectroscopy and positron emission tomography, but here, we discuss some of the wide-field TCSPC methods for applications in fluorescence microscopy. We describe work by us and others as presented in the Ulitima fast imaging and tracking conference at the Argonne National Laboratory in September 2018, from phosphorescence lifetime imaging (PLIM) microscopy on the microsecond time scale to fluorescence lifetime imaging (FLIM) on the nanosecond time scale, and highlight some applications of these techniques

# Wide-field Time-correlated Single Photon Counting-based Fluorescence Lifetime Imaging Microscopy

Klaus Suhling<sup>\*a</sup>, Liisa M. Hirvonen<sup>a,h</sup>, Wolfgang Becker<sup>b</sup>, Stefan Smietana<sup>b</sup>, Holger Netz<sup>b</sup>, James Milnes<sup>c</sup>, Thomas Conneely<sup>c</sup>, Alix Le Marois<sup>a,i</sup>, Ottmar Jagutzki<sup>d</sup>, Fred Festy<sup>e,j</sup>, Zdeněk Petrášek<sup>f</sup>, Andrew Beeby<sup>g</sup>

<sup>a</sup>Department of Physics, King's College London, Strand, London WC2R 2LS, UK

<sup>b</sup>Becker & Hickl GmbH, Nahmitzer Damm 30, 12277 Berlin, Germany

<sup>c</sup>Photek Ltd, 26 Castleham Rd, St Leonards on Sea TN38 9NS, UK

<sup>d</sup>Institut für Kernphysik, Max-von-Laue-Str. 1, 60438 Frankfurt, Germany

<sup>e</sup>Biomaterials, Biomimetics and Biophotonics Research Group, Kings College London Dental Institute at Guys Hospital, Kings Health Partners, Guys Dental Hospital, London Bridge, London SE1 9RT, UK

<sup>f</sup>Institut für Biotechnologie und Bioprozesstechnik, Technische Universität Graz, Petersgasse 10-12/I, 8010 Graz, Austria

<sup>g</sup>Department of Chemistry, University of Durham, Durham DH13LE, UK

<sup>h</sup>present address: Randall Centre for Cell & Molecular Biophysics, New Hunt's House, King's College London, Guy's Campus, London SE1 1UL, UK

<sup>i</sup>present address: The Francis Crick Institute, 1 Midland Road, London NW1 1AT, UK

<sup>j</sup>present address: Virtual AI, 24 Oakwood, Berkhamsted, Hertfordshire HP4 3NQ, UK

## Abstract

Wide-field time-correlated single photon counting detection techniques, where the position and the arrival time of the photons are recorded simultaneously using a camera, have made some advances recently. The technology and instrumentation used for this approach is employed in areas such as nuclear science, mass spectroscopy and positron emission tomography, but here, we discuss some of the wide-field TCSPC methods for applications in fluorescence microscopy. We describe work by us and others as presented in the Ulitima fast imaging and tracking conference at the Argonne National Laboratory in September 2018, from phosphorescence lifetime imaging (PLIM) microscopy on the microsecond time scale to fluorescence lifetime imaging (FLIM) on the nanosecond time scale, and highlight some applications of these techniques.

## 1 Introduction

Optical microscopy is a widely-used tool for non-destructive and minimally invasive observation of living samples. Fluorescence microscopy in particular allows the observation of cell dynamics and function in real time with negligible cytotoxicity. It can also provide high sensitivity, down to the single molecule level, and high specificity. In addition to localising fluorescent labels, the fluorescence can also be used for sensing the immediate environment of the fluorophore, via its spectral properties, its polarization or fluorescence lifetime.[1] The fluorescence lifetime is often used for this, as it is independent of the concentration of the fluorescence probe, which is difficult to control in cells.

The fluorescence lifetime is the average time a fluorophore remains in the excited state, typically nanoseconds. It can be a function of viscosity, temperature, pH, ion or glucose concentration, refractive index or polarity, and of interaction with other molecules, e.g. due to Förster Resonance Energy Transfer (FRET), a widely used technique to identify protein conformational changes or interactions. Phosphorescence, which originates from the probe's triplet state, is typically on the microsecond time scale, and is used to sense oxygen concentration, with phosphorescence lifetime imaging (PLIM) being the imaging version of this approach. [2, 3] Fluorescence lifetime imaging (FLIM) and PLIM can measure the lifetimes in every pixel, thus providing image contrast according to the lifetime, which then provides contrast according to viscosity, oxygen or ion concentration or temperature, depending on the type of fluorescent probe employed.

FLIM or PLIM is often carried out with time-correlated single photon counting (TCSPC) and confocal or multiphoton beam scanning. TCSPC is effectively a delayed coincidence method,[4] the origin of which lies in particle physics. It can be traced back to 1929, when Bothe and Kohlhörster used two Geiger counters separated by spacers of varying thickness to study coincidences of penetrating charged particles in cosmic rays.[5] This was followed by the first practical electronic coincidence circuit by Rossi in 1930, which became a precursor of the AND logic gate in electronic circuits.[6] By the addition of a delay, this coincidence method evolved to measure delayed coincidence, and thus provided the means for time-resolved measurements. A method to measure the amplitude of the signal as a function of the delay between pulses by Rossi in 1942, a "time-circuit", is now known

as the time-to-amplitude converter (TAC).[6] It became a popular method to measure short radioactive decays in the 1950s,[7] and in 1961, Bollinger and Thomas generalised the scintillation measurements enabling TCSPC as we now know it, where the arrival time of a single photon is measured relative to an excitation pulse.[8] The accumulation of many single photons then represents the intensity decay of the sample, as long as no photons are lost due to pile-up, and the linearity between intensity and collected photons holds. The first reports that use TCSPC in the measurement of fluorescence decays appear in the early 1970s,[9-11] and TCSPC was soon widely used for time-resolved spectroscopy, and in particular the measurement of fluorescence lifetimes in solutions. Flashlamps used kHz repetition rates,[12] but lasers, with picosecond excitation pulses at MHz repetition rates, sped up the measurements significantly and advanced this field enormously.[13] In addition to scintillation and fluorescence measurements, TCSPC is also used for lidar [14] and optical tomography.[15]


TCSPC is a sensitive, precise, robust and mature technique to measure photon arrival times after an excitation pulse.[4] Its advantages stem from the digital nature of the technique, based on whether a single photon is detected, or not. It obeys well-defined Poisson statistics, which state that the experimental uncertainty is the square root of the number of counts, and that the signal-to-noise ratio increases with the measurement time, i.e. the number of counts. It also has a large dynamic range in time, from picoseconds to microseconds, and affords an easy visualisation of fluorescence decays. TCSPC has the highest signal-to-noise ratio of the standard time-resolved imaging methods,[16-18] and is precise enough to permit multi-exponential fluorescence lifetime fitting. By raster scanning the focal spot over the sample the image is created pixel by pixel, using single point detectors to perform TCSPC in each pixel.

While TCSPC is straight-forward to implement with scanning microscopy, there are a number of fluorescence microscopy methods that employ a camera, for example lightsheet microscopy, total internal reflection fluorescence (TIRF), supercritical angle fluorescence and super-resolution fluorescence microscopy methods based on localisation of fluorophores. To harness the advantages of TCSPC for these camera-based fluorescence microscopy methods, and to perform single photon sensitive wide-field FLIM or PLIM with these microscopy methods, a single photon sensitive camera with appropriate time resolution is required.[19-22] Sensitive detection is favoured, as fluorophores eventually bleach.[23] For microsecond time resolution PLIM, microchannel plate (MCP)-based photon counting image intensifiers with a phosphor screen and a fast camera can be used. For picosecond resolution FLIM, special read-out schemes for image intensifiers, e.g. charge division and propagation time techniques can be employed.

In addition to visible light, single photon detectors based on optoelectronic vacuum devices such as image intensifiers can also detect other types of electromagnetic radiation, including UV photons, x-rays and gamma rays, as well as particles, such as electrons, neutrons and ion fragments. MCPs are used in time-of-flight mass spectroscopy for detecting ion fragments, where molecules in vacuum chamber are ionised and broken up by short laser pulses.[24, 25] The fragments are accelerated by an electric field towards the MCP, and the position and arrival time contains information about the molecular fragments created. Furthermore, boron-doped MCPs are used for neutron detection and imaging,[26-28] as boron has a high neutron absorption cross section.[29] Timing of the neutron events allows distinction of cold and thermal neutrons, [27] and thus enables simultaneous imaging at different neutron energies. This yields different contrast, as absorption cross sections of the material in the object under study are a function of neutron energy. Moreover, in autoradiography, where radioactively labelled samples are imaged, MCPs can be used for sensitive detection of weak  $\beta$ -emitters such as  $^3\text{H}$ ,  $^{14}\text{C}$ , and  $^{35}\text{S}$ . [30] We note that for these approaches, once a particle has been converted into an electron in the MCP, the detection process is the same as for photons. Thus, development of wide-field detectors for FLIM and PLIM may also benefit these applications.

## **2. Wide-field time-correlated single photon counting imaging techniques and applications**


Photon counting imaging, where the image is assembled from individual photons, effectively allows read-out noise free imaging. The approach goes back to the 1980s [31] and was, for example, used on the Faint Object Camera on the Hubble Space Telescope.[32] Note that the prime consideration at the time was sensitivity, not photon arrival timing. Modern MCP-based image intensifiers operated at saturated gain can perform photon counting imaging, and can either be equipped with a phosphor screen output, to be viewed with a camera, or with a position-sensitive anode.



## 2.1 Image intensifiers with a phosphor screen and camera readout for microsecond decay measurements

### 2.1.1 Photon arrival time obtained directly from the camera frame rate

Recent developments in imaging sensor technology have allowed complementary metal oxide semiconductor (CMOS) cameras to reach MHz frame rates.[33] These cameras can be used in combination with a photon counting image intensifier for TCSPC.[34] After each excitation pulse, a sequence of frames is acquired during the decay time of the probe, and this process is repeated until enough photons are collected so that a decay histogram is obtained for each pixel of the image, as shown in Figure 1. The time resolution of this approach is limited by the camera frame rate to the microsecond time scale. Although there is a trade-off between a high frame rate and the number of pixels that can be imaged, this technique enables the collection of up to hundreds of photons per frame, and even several photons after one excitation cycle per pixel as long as they arrive in different frames.[34] Latest developments of sensors that detect a signal above a certain threshold, e.g. timepix[35, 36] or pimms[37] cameras, can also be used for this purpose.[38] The details of the photon event intensity and area covered (number of camera pixels) is of no importance, only the fact where and when a photon event occurs, and therefore such an approach is feasible.




### 2.1.2 Photon arrival time obtained from imaging the phosphor decay

The phosphor decay of the intensifier output screen can be exploited to find the photon arrival time within the frame exposure time.[39] This effect, an afterglow, is usually undesired,[40] but can be put to good use. Matching the camera frame rate to the phosphor decay such that the photon events can be seen in several consecutive frames allows the photon arrival time within the camera exposure time to be found from the relative brightness of the photon event in successive frames, as illustrated in Figure 2. This approach is similar to dual exposure techniques for velocity map imaging in mass spectroscopy.[41, 42] The measurement of the photon arrival time from the phosphor decay can improve the time resolution beyond the inverse frame rate of the camera, and the lower frame rate increases the number of recorded pixels, thus allowing bigger field of view.[43, 44] Phosphorescent sample decays as short as 500 ns have been measured with a P20 phosphor and 300 kHz frame rate [45], as shown in Figure 3, but a combination of a faster phosphor and a faster frame rate, or special cameras such as could allow the measurement of even faster sample decays.

Camera-based wide-field TCSPC is especially well suited for sensitive measurements of phosphorescence lifetimes in the micro- and millisecond time region, i.e. PLIM.[2, 3] Camera-based methods enable the collection of hundreds of photons per excitation cycle, shortening the data acquisition time with long lifetime probes compared to single point scanning measurements. Lifetimes around 1  $\mu$ s have been measured with several transition metal probes using these techniques, including a ruthenium-based oxygen sensor in living cells, with total image acquisition times of just a few seconds.[33, 43] In fact, the best use of this approach is probably for



microsecond lifetime PLIM measurements due to the limited overall count rate of the image intensifier. It enables the collection of many photons per excitation pulse with low excitation power and without lengthy scanning.




## 2.2 Image intensifiers with an electronic anode readout for nanosecond decay measurements

For nanosecond fluorescence lifetime measurements, tens or hundreds of picosecond precision for photon arrival timing is essential. While the MCPs themselves are capable of timing the photon arrival with a precision of a few tens of picoseconds,[4] they are not capable of recording the photon's position without special read-outs. Different read-out architectures have been developed, including quadrant, wedge-and-strip, cross-strip and delay line anodes,[19, 20] where the position of the electron cloud is determined via a charge division approach, or via the propagation time along a delay line.[46] Although some of the read-out schemes can accommodate multiple photon events after one excitation cycle, i.e. the hexanode [47] or cross-strip read-outs,[19] the count rate is typically limited by the position readout electronics to a few 10s or 100s of kHz, rather than the need to avoid overlapping events. However, the advantage here is that sub-microwatt excitation powers are sufficient to generate photons, and allow continuous observation of living samples over days. Wide-field data collection allows the tracking of individual molecules or particle trajectories, e.g. single quantum dot tracking and FLIM has been demonstrated with a MCP and delay line anode detector.[48] Quadrant anodes [49, 50] have been applied to FLIM for the study of protein-protein interaction by FRET [51-53] and photosynthesis.[54, 55]

Wide-field TCSPC is especially useful for microscopy methods where the whole field of view is illuminated with a technique that provides depth discrimination. One of these techniques is TIRF microscopy, where the sample is excited by an evanescent wave only near (up to 100 nm) the coverslip. Total internal reflection (TIR)-FLIM has been demonstrated with quadrant anodes [56] and SPAD detectors.[57]

We have used a delay line anode detector (Photek) where the delay line is capacitively coupled to a resistive anode inside the tube and using an image charge technique,[20, 58] to perform TIR-FLIM.[59] Figure 4 shows wide-field TCSPC FLIM images of fixed HeLa cells acquired with this delay line anode detector, read out by conventional TCSPC timing boards (Becker & Hickl).[60] The sample was excited with a Horiba DeltaDiode picosecond laser (375 nm) at 10 MHz, and the photon count rate was around 80 kHz, with an acquisition time of 100s. The measured intensity shows the cell membrane stained with membrane dye laurdan only under TIRF illumination (Figure 4b), while the whole cell is visible under wide-field illumination (Figure 4c). The laurdan fluorescence lifetime is also shortened under TIRF illumination (Figure 4d) compared to wide-field illumination (Figure 4f); one contributory factor here is the proximity of the high refractive index glass coverslip which consequently shortens the fluorescence lifetime.[61] Figure 4g shows the measured instrument response function (IRF), and the measured time decays in a small area in Figure 4c,d. The IRF full width at half maximum (FWHM) is 344 ps.

The low excitation power used in wide-field TCSPC can help to minimise photodamage in living cells which is especially beneficial for observing dynamics in living cells over long periods.



### 2.3 Electron-bombarded sensors

Single photon detection is also possible with electron-bombarded (EB) sensors, where the photoelectrons from the photocathode are accelerated directly into a CCD or CMOS sensor.[62-66] Unlike MCPs where the statistical electron multiplication process creates a broad pulse height distribution, in EB sensors the photon event brightness depends on the gain voltage. Thus, by sweeping the gain voltage during the exposure time, it could be possible to obtain photon arrival time information from the photon event brightness, akin to a 2-dimensional streak camera.[62] The concept has been proposed and simulated, but not implemented; gain sweeping has not been possible with commercially available devices.

### 2.4 Single photon avalanche diode arrays

Another option to obtain a position-sensitive single photon detector is to build an array of single photon sensitive point detectors.[67-69] A single photon avalanche diode (SPAD, reverse biased above the diode breakdown voltage, Geiger mode) is a small all-solid state photon detector with a diameter of a few microns, and is capable of single photon detection with picosecond time resolution.[70] Single SPADs were first used for fast timing applications in the 1980s, and the implementation of SPADs in CMOS technology in 2003 enabled the development of SPAD arrays.[70] Unlike image intensifiers or EB sensors, SPADs do not require a high voltage or a vacuum, they are not damaged by high light levels, and they can be manufactured in arrays, 256×256 pixel [71] and 240×320 pixel SPAD arrays [72] have been reported.

SPAD arrays are a relatively new development in wide-field TCSPC. Initially, gated SPAD arrays for fluorescence lifetime measurements were implemented [73], but now each photon can be timed individually, and in all pixels in parallel.[70] The big advantage of these developments in SPAD array detector technology is that it allows independent TCSPC in each pixel of a SPAD array simultaneously, e.g. in the 32 × 32 pixel megafame chip, with a TDC in each pixel, with 55 ps resolution.[74, 75] They simultaneously deliver single photon sensitivity, tens of thousands of pixels spatial resolution and picosecond timing resolution.[70, 76, 77] The outstanding capability of enormous global count rates well into the gigahertz region,[78] which would allow the observation of fast cellular dynamics, is a big advantage of these devices.



The design of detectors and timing electronics on a single substrate inevitably provides compactness and large numbers of channels but compromises fill-factor and SPAD performance (jitter, photon detection efficiency, after pulsing and dark count). Nevertheless, SPAD array technology offers a huge advantage over existing FLIM detector technology. SPAD array detectors currently have a small fill factor (<10%), because the majority of the area of each pixel is occupied by electronic circuits to perform the timing, with only a small light-sensitive area dedicated to the detection of photons. Promising current developments in 3D stacking of integrated circuits [79] will ensure a fill factor >80%, a better time resolution and reduced jitter. Moreover, logic integration will scale up enormously enabling placement of field programmable gate arrays-like structures beneath the sensor. The development of 100% fill factor SPAD arrays will not only allow fast fluorescence lifetime measurements via wide-field TCSPC FLIM, the sensitivity and speed of this kind of detector could also benefit other applications, and this field continues to develop at a fast pace.[70, 77]. SPAD arrays are also used for positron emission tomography, where the picosecond timing capabilities can pinpoint the localization of the annihilation event more precisely than without timing, and thus increase the spatial resolution of the technique.[80] In addition, SPAD arrays have been used for detection of Cherenkov radiation in radiation therapy.[81] Range-finding is another application of SPAD arrays,[14] for example consumer electronics such as mobile phones have benefitted from TCSPC-based range finding, where sensitivity is not as big an issue as it is with FLIM.<sup>1</sup>

### 3. Conclusions

Fluorescence microscopy allows non-destructive and minimally invasive observation of living samples, and FLIM and PLIM allow the monitoring of the microenvironment of fluorescence and phosphorescence probes. A photon counting approach to FLIM and PLIM is particularly helpful, as it minimizes light exposure of the sample. It is also the most sensitive method to collect the fluorescence from the sample before the fluorophores are irreversibly bleached.[23] Wide-field TCSPC-based methods combine the advantages of single photon sensitivity and precision with wide-field data collection. This is important for implementation of specialised FLIM and PLIM microscopy methods that typically employ cameras, such as TIRF, lightsheet and others.

We have shown that wide-field time-correlated single photon counting based on an image intensifier with a phosphor screen and a fast CMOS camera can be employed for PLIM to map phosphorescent decays on a microsecond time scale. This can either be done with direct imaging of the photon events on the image intensifier's phosphor screen,[33] or by exploiting the invariant phosphor decay of the image intensifier screen for accurate timing of photon arrival well below the camera exposure time.[43] To image nanosecond fluorescence decays, we show that a crossed delay line anode detector read out with conventional TCSPC boards is feasible.[59] This approach retains all the advantages of TCSPC, and extends them to wide-field detection, serving essentially as a single photon sensitive camera with picosecond time resolution. Application of this approach to TIR FLIM using the membrane dye laurdan shows lifetime contrast between the plasma membrane and the interior membranes of the cell, as shown in figure 4. Photon counting approaches are particularly useful for fluorescence microscopy methods employing a camera to enable TCSPC-based FLIM, the FLIM method with the highest signal-to-noise ratio. The extremely low illumination intensity, distributed evenly over the field of view, is beneficial especially in life science applications where it allows long-term monitoring of living cells and organisms, while wide-field data collection enables the observation of cell dynamics and single particle tracking. A position-sensitive photon counting detector has recently been used for simultaneous acquisition of both spectral and temporal information of Raman photons from tissue phantoms,[82] showing the versatility of these types of detector.

Wide-field TCSPC methods are currently mainly based on MCP-based detectors, a mature technology used especially in astronomy and medical imaging. However, the development of SPAD arrays and their application to FLIM with the prospect of huge photon count rates continues at a rapid pace.

### 4. Acknowledgements

KS and LMH acknowledge BBSRC Sparking Impact and EPSRC Accelerated Impact funding for this work, as well as BBSRC grant BB/R004803.

---

<sup>1</sup> See [http://www.st.com/content/st\\_com/en/about/media-center/press-item.html/stmicroelectronics-proximity-sensor-solves-smartphone-hang-ups.html](http://www.st.com/content/st_com/en/about/media-center/press-item.html/stmicroelectronics-proximity-sensor-solves-smartphone-hang-ups.html) for information about proximity sensing in mobile phones using photon time-of-flight measurements with SPAD arrays.

## References

1. Suhling, K., L.M. Hirvonen, J.A. Levitt, P.-H. Chung, C. Tregidgo, A. Le Marois, D.A. Rusakov, K. Zheng, S. Ameer-Beg, S. Poland, S. Coelho, R. Henderson, and N. Krstajic, *Fluorescence lifetime imaging (FLIM): Basic concepts and some recent developments*. Medical Photonics, 2015. **27**: p. 3-40.
2. Shcheslavskiy, V.I., A. Neubauer, R. Bukowiecki, F. Dinter, and W. Becker, *Combined fluorescence and phosphorescence lifetime imaging*. Applied Physics Letters, 2016. **108**(9): p. 091111.
3. Baggaley, E., S.W. Botchway, J.W. Haycock, H. Morris, I.V. Sazanovich, J.A.G. Williams, and J.A. Weinstein, *Long-lived metal complexes open up microsecond lifetime imaging microscopy under multiphoton excitation: from FLIM to PLIM and beyond*. Chemical Science, 2014. **5**(3): p. 879-886.
4. O'Connor, D.V. and D. Phillips, *Time-correlated single-photon counting*. 1984, London: Academic Press.
5. Bothe, W. and W. Kolhörster, *Das Wesen der Höhenstrahlung*. Zeitschrift für Physik, 1929. **56**(11): p. 751-777.
6. Rossi, B., *Method of Registering Multiple Simultaneous Impulses of Several Geiger's Counters*. Nature, 1930. **125**: p. 636.
7. Bell, R.E., R.L. Graham, and H.E. Petch, *Design and use of a coincidence circuit of short resolving time*. Canadian Journal of Physics, 1952. **30**(1): p. 35-52.
8. Bollinger, L.M. and G.E. Thomas, *Measurement of the Time Dependence of Scintillation Intensity by a Delayed-Coincidence Method*. Review of Scientific Instruments, 1961. **32**(9): p. 1044-1050.
9. Binkert, T., H.P. Tschanz, and P.E. Zinsli, *The measurement of fluorescence decay curves with the single-photon counting method and the evaluation of rate parameters*. Journal of Luminescence, 1972. **5**(3): p. 187-217.
10. Lewis, C., W.R. Ware, L.J. Doemeny, and T.L. Nemzek, *The Measurement of Short-Lived Fluorescence Decay Using the Single Photon Counting Method*. Review of Scientific Instruments, 1973. **44**(2): p. 107-114.
11. Bachrach, R.Z., *A Photon Counting Apparatus for Kinetic and Spectral Measurements*. Review of Scientific Instruments, 1972. **43**(5): p. 734-737.
12. Birch, D.J.S. and R.E. Imhof, *Coaxial nanosecond flashlamp*. Review of Scientific Instruments, 1981. **52**(8): p. 1206-1212.
13. Phillips, D., *A lifetime in photochemistry: some ultrafast measurements on singlet states*. Proceedings of the Royal Society A: Mathematical, Physical and Engineering Sciences, 2016. **472**(2190): p. 20160102.
14. McCarthy, A., R.J. Collins, N.J. Krichel, V. Fernandez, A.M. Wallace, and G.S. Buller, *Long-range time-of-flight scanning sensor based on high-speed time-correlated single-photon counting*. Applied Optics, 2009. **48**(32): p. 6241-6251.
15. Schmidt, F.E.W., M.E. Fry, E.M.C. Hillman, J.C. Hebden, and D.T. Delpy, *A 32-channel time-resolved instrument for medical optical tomography*. Review of Scientific Instruments, 2000. **71**(1): p. 256-265.
16. Esposito, A., H.C. Gerritsen, and F.S. Wouters, *Optimizing frequency-domain fluorescence lifetime sensing for high-throughput applications: photon economy and acquisition speed*. Journal of the Optical Society of America A, 2007. **24**: p. 3261-3273.
17. Gratton, E., S. Breusegem, J. Sutin, Q. Ruan, and N. Barry, *Fluorescence lifetime imaging for the two-photon microscope: time-domain and frequency-domain methods*. Journal of Biomedical Optics, 2003. **8**(3): p. 381-390.
18. Philip, J. and K. Carlsson, *Theoretical investigation of the signal-to-noise ratio in fluorescence lifetime imaging*. Journal of the Optical Society of America A, 2003. **20**(2): p. 368-379.
19. Michalet, X., R.A. Colyer, G. Scalia, A. Ingargiola, R. Lin, J.E. Millaud, S. Weiss, O.H.W. Siegmund, A.S. Tremsin, J.V. Vallerger, A. Cheng, M. Levi, D. Aharoni, K. Arisaka, F. Villa, F. Guerrieri, F. Panzeri, I. Rech, A. Gulinatti, F. Zappa, M. Ghioni, and S. Cova, *Development of new photon-counting detectors for single-molecule fluorescence microscopy*. Philosophical Transactions of the Royal Society B-Biological Sciences, 2013. **368**(1611): p. 20120035.
20. Jagutzki, O., J.S. Lapington, L.B.C. Worth, U. Spillman, V. Mergel, and H. Schmidt-Böcking, *Position sensitive anodes for MCP read-out using induced charge measurement*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2002. **477**(1-3): p. 256-261.
21. Hartig, R., Y. Prokazov, E. Turbin, and W. Zusratter, *Wide-field fluorescence lifetime imaging with multi-anode detectors*, in *Fluorescence Spectroscopy and Microscopy: Methods and Protocols, Methods in Molecular Biology*, Y. Engelborghs and A.J.W.G. Visser, Editors. 2014, Springer. p. 457-480.
22. Prokazov, Y., E. Turbin, A. Weber, R. Hartig, and W. Zusratter, *Position sensitive detector for fluorescence lifetime imaging*. Journal of Instrumentation, 2014. **9**: p. C12015.

23. Zhao, Q.L., I.T. Young, and J.G.S. de Jong, *Photon budget analysis for fluorescence lifetime imaging microscopy*. Journal of Biomedical Optics, 2011. **16**(8): p. 086007.
24. Vallance, C., M. Brouard, A. Lauer, C.S. Slater, E. Halford, B. Winter, S.J. King, J.W.L. Lee, D.E. Pooley, I. Sedgwick, R. Turchetta, A. Nomerotski, J.J. John, and L. Hill, *Fast sensors for time-of-flight imaging applications*. Physical Chemistry Chemical Physics, 2014. **16**(2): p. 383-395.
25. Jungmann, J.H. and R.M.A. Heeren, *Detection systems for mass spectrometry imaging: A perspective on novel developments with a focus on active pixel detectors*. Rapid Communications in Mass Spectrometry, 2013. **27**(1): p. 1-23.
26. Tremsin, A.S., J.B. McPhate, J.V. Vallerga, O.H.W. Siegmund, W. Bruce Feller, E. Lehmann, A. Kaestner, P. Boillat, T. Panzner, and U. Filges, *Neutron radiography with sub-15  $\mu\text{m}$  resolution through event centroiding*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2012. **688**: p. 32-40.
27. Tremsin, A.S., J.B. McPhate, J.V. Vallerga, O.H.W. Siegmund, J.S. Hull, W.B. Feller, and E. Lehmann, *Detection efficiency, spatial and timing resolution of thermal and cold neutron counting MCP detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2009. **604**(1): p. 140-143.
28. Siegmund, O.H., J.V. Vallerga, A.S. Tremsin, J. McPhate, and B. Feller, *High spatial resolution neutron sensing microchannel plate detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2007. **576**(1): p. 178-182.
29. Fraser, G.W. and J.F. Pearson, *The direct detection of thermal neutrons by imaging microchannel-plate detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 1990. **293**(3): p. 569-574.
30. Tarhoni, M.H., V. Vigneswara, M. Smith, S. Anderson, P. Wigmore, J.E. Lees, D.E. Ray, and W.G. Carter, *Detection, Quantification, and Microlocalisation of Targets of Pesticides Using Microchannel Plate Autoradiographic Imagers*. Molecules, 2011. **16**(10): p. 8535-8551.
31. Boksenberg, A., *Advances in Detectors for Astronomical Spectroscopy*. Philosophical Transactions of the Royal Society of London Series a-Mathematical Physical and Engineering Sciences, 1982. **307**(1500): p. 531-548.
32. Kröger, H.W., G.K. Schmidt, and N. Pailer, *Faint object camera: European contribution to the Hubble Space Telescope*. Acta Astronautica, 1992. **26**(11): p. 827-834.
33. Hirvonen, L.M., F. Festy, and K. Suhling, *Wide-field time-correlated single-photon counting (TCSPC) lifetime microscopy with microsecond time resolution*. Optics Letters, 2014. **39**(19): p. 5602-5605.
34. Sergeant, N., J.A. Levitt, M. Green, and K. Suhling, *Rapid wide-field photon counting imaging with microsecond time resolution*. Optics Express, 2010. **18**(24): p. 25292-25298.
35. Fisher-Levine, M. and A. Nomerotski, *TimepixCam: a fast optical imager with time-stamping*. Journal of Instrumentation, 2016. **11**(03): p. C03016.
36. Zhao, A., M. van Beuzekom, B. Bouwens, D. Byelov, I. Chakaberia, C. Cheng, E. Maddox, A. Nomerotski, P. Svihra, J. Visser, V. Vrba, and T. Weinacht, *Coincidence velocity map imaging using Tpx3Cam, a time stamping optical camera with 1.5 ns timing resolution*. Review of Scientific Instruments, 2017. **88**(11): p. 113104.
37. John, J.J., M. Brouard, A. Clark, J. Crooks, E. Halford, L. Hill, J.W.L. Lee, A. Nomerotski, R. Pisarczyk, I. Sedgwick, C.S. Slater, R. Turchetta, C. Vallance, E. Wilman, B. Winter, and W.H. Yuen, *PImMS, a fast event-triggered monolithic pixel detector with storage of multiple timestamps*. Journal of Instrumentation, 2012. **7**(08): p. C08001.
38. Hirvonen, L.M., M. Fisher-Levine, K. Suhling, and A. Nomerotski, *Photon counting phosphorescence lifetime imaging with TimepixCam*. Review of Scientific Instruments, 2017. **88**(1): p. 013104.
39. Petrášek, Z. and K. Suhling, *Photon arrival timing with sub-camera exposure time resolution in wide-field time-resolved photon counting imaging*. Optics Express, 2010. **18**(24): p. 24888-24901.
40. Mainprize, J.G. and M.J. Yaffe, *The effect of phosphor persistence on image quality in digital x-ray scanning systems*. Medical Physics, 1998. **25**(12): p. 2440-2454.
41. Dinu, L., A.T.J.B. Eppink, F. Rosca-Pruna, H.L. Offerhaus, W.J. van der Zande, and M.J.J. Vrakking, *Application of a time-resolved event counting technique in velocity map imaging*. Review of Scientific Instruments, 2002. **73**(12): p. 4206-4213.
42. Strasser, D., X. Urbain, H.B. Pedersen, N. Altstein, O. Heber, R. Wester, K.G. Bhushan, and D. Zajfman, *An innovative approach to multiparticle three-dimensional imaging*. Review of Scientific Instruments, 2000. **71**(8): p. 3092-3098.
43. Hirvonen, L.M., Z. Petrášek, A. Beeby, and K. Suhling, *Sub- $\mu\text{s}$  time resolution in wide-field time-correlated single photon counting microscopy obtained from the photon event phosphor decay*. New Journal of Physics, 2015. **17**(2): p. 023032.

44. Hirvonen, L.M., Z. Petrášek, and K. Suhling, *Wide-field time-correlated single photon counting (TCSPC) microscopy with time resolution below the frame exposure time*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2015. **787**: p. 1-5.
45. Hirvonen, L.M., Z. Petrášek, A. Beeby, and K. Suhling, *Microsecond wide-field TCSPC microscopy based on an ultra-fast CMOS camera*. 2015, SPIE Proc 9329. p. 932939.
46. Hirvonen, L.M. and K. Suhling, *Wide-field TCSPC: methods and applications*. Measurement Science and Technology, 2017. **28**(1): p. 012003.
47. Jagutzki, O., A. Cerezo, A. Czasch, R. Dörner, M. Hattass, M. Huang, V. Mergel, U. Spillmann, K. Ullmann-Pfleger, T. Weber, H. Schmidt-Böcking, and G.D.W. Smith, *Multiple hit readout of a microchannel plate detector with a three-layer delay-line anode*. IEEE Transactions on Nuclear Science, 2002. **49**(5): p. 2477-2483.
48. Michalet, X., R.A. Colyer, J. Antelman, O.H.W. Siegmund, A. Tremsin, J.V. Vallerga, and S. Weiss, *Single-Quantum Dot Imaging with a Photon Counting Camera*. Current Pharmaceutical Biotechnology, 2009. **10**: p. 543-558.
49. Kemnitz, K., L. Pfeifer, and M.R. Ainbund, *Detector for multichannel spectroscopy and fluorescence lifetime imaging on the picosecond timescale*. Nuclear Instruments and Methods in Physics Research Section A, 1997. **387**(1/2): p. 86-87.
50. Emiliani, V., D. Sanvitto, M. Tramier, T. Piolot, Z. Petrášek, K. Kemnitz, C. Duneux, and M. Coppey-Moisán, *Low-intensity two-dimensional imaging of fluorescence lifetimes in living cells*. Applied Physics Letters, 2003. **83**(12): p. 2471-2473.
51. Vitali, M., F. Picazo, Y. Prokazov, A. Duci, E. Turbin, C. Gotze, J. Llopis, R. Hartig, A.J.W.G. Visser, and W. Zschratte, *Wide-Field Multi-Parameter FLIM: Long-Term Minimal Invasive Observation of Proteins in Living Cells*. Plos One, 2011. **6**(2): p. e15820.
52. Ramadass, R., D. Becker, M. Jendrach, and J. Bereiter-Hahn, *Spectrally and spatially resolved fluorescence lifetime imaging in living cells: TRPV4-microfilament interactions*. Archives of Biochemistry and Biophysics, 2007. **463**(1): p. 27-36.
53. Gautier, I., M. Tramier, C. Durieux, J. Coppey, R.B. Pansu, J.C. Nicolas, K. Kemnitz, and M. Coppey-Moisán, *Homo-FRET microscopy in living cells to measure monomer-dimer transition of GFP-tagged proteins*. Biophysical Journal, 2001. **80**(6): p. 3000-3008.
54. Petrášek, Z., H.J. Eckert, and K. Kemnitz, *Wide-field photon counting fluorescence lifetime imaging microscopy: application to photosynthesizing systems*. Photosynthesis Research, 2009. **102**(2-3): p. 157-168.
55. Petrášek, Z., F.J. Schmitt, C. Theiss, J. Huyer, M. Chen, A. Larkum, H.J. Eichler, K. Kemnitz, and H.J. Eckert, *Excitation energy transfer from phycobiliprotein to chlorophyll d in intact cells of Acaryochloris marina studied by time-and wavelength-resolved fluorescence spectroscopy*. Photochemical & Photobiological Sciences, 2005. **4**(12): p. 1016-1022.
56. Giraud, G., H. Schulze, T. Bachmann, C. Campbell, A. Mount, P. Ghazal, M. Khondoker, A. Ross, S. Ember, I. Ciani, C. Tlili, A. Walton, J. Terry, and J. Crain, *Fluorescence Lifetime Imaging of Quantum Dot Labeled DNA Microarrays*. International Journal of Molecular Sciences, 2009. **10**(4): p. 1930-1941.
57. Giraud, G., H. Schulze, D.U. Li, T.T. Bachmann, J. Crain, D. Tyndall, J. Richardson, R. Walker, D. Stoppa, E. Charbon, R. Henderson, and J. Arlt, *Fluorescence lifetime biosensing with DNA microarrays and a CMOS-SPAD imager*. Biomedical Optics Express, 2010. **1**(5): p. 1302-1308.
58. Milnes, J., J.S. Lapington, O. Jagutzki, and J. Howorth, *Image charge multi-role and function detectors*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2009. **604**(1-2): p. 218-220.
59. Hirvonen, L.M., W. Becker, J. Milnes, T. Conneely, S. Smietana, A. Le Marois, O. Jagutzki, and K. Suhling, *Picosecond wide-field time-correlated single photon counting fluorescence microscopy with a delay line anode detector*. Applied Physics Letters, 2016. **109**: p. 071101.
60. Becker, W., L.M. Hirvonen, J. Milnes, T. Conneely, O. Jagutzki, H. Netz, S. Smietana, and K. Suhling, *A wide-field TCSPC FLIM system based on an MCP PMT with a delay-line anode*. Review of Scientific Instruments, 2016. **87**: p. 093710.
61. Tregidgo, C., J.A. Levitt, and K. Suhling, *Effect of refractive index on the fluorescence lifetime of green fluorescent protein*. Journal of Biomedical Optics, 2008. **13**: p. 031218.
62. Hirvonen, L.M., S. Jiggins, N. Sergeant, G. Zanda, and K. Suhling, *Photon counting imaging with an electron-bombarded CCD: towards a parallel-processing photoelectronic time-to-amplitude converter*. Rev Sci Instrum, 2014. **85**(12): p. 123102.
63. Hirvonen, L.M., S. Jiggins, N. Sergeant, G. Zanda, and K. Suhling, *Photon counting imaging with an electron-bombarded CCD: Towards wide-field time-correlated single photon counting (TCSPC)*.

- Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2015. **787**: p. 323-327.
64. Cajgfinger, T., A. Dominjon, and R. Barbier, *Single photon detection and localization accuracy with an ebCMOS camera*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2015. **787**: p. 176-181.
65. Buontempo, S., et al., *The Megapixel EBCCD: A high-resolution imaging tube sensitive to single photons*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 1998. **413**(2-3): p. 255-262.
66. Hirvonen, L.M. and K. Suhling, *Photon Counting Imaging with an Electron-Bombarded Pixel Image Sensor*. Sensors, 2016. **16**(5): p. 617.
67. Rinnenthal, J.L., C. Börnchen, H. Radbruch, V. Andresen, A. Mossakowski, V. Siffrin, T. Seelemann, H. Spiecker, I. Moll, J. Herz, A.E. Hauser, F. Zipp, M.J. Behne, and R. Niesner, *Parallelized TCSPC for Dynamic Intravital Fluorescence Lifetime Imaging: Quantifying Neuronal Dysfunction in Neuroinflammation*. Plos One, 2013. **8**(4): p. e60100.
68. McLoskey, D., D.J.S. Birch, A. Sanderson, K. Suhling, E. Welch, and P.J. Hicks, *Multiplexed single-photon counting. I. A time-correlated fluorescence lifetime camera*. Review of Scientific Instruments, 1996. **67**(6): p. 2228-2237.
69. Suhling, K., D. McLoskey, and D.J.S. Birch, *Multiplexed single-photon counting. II. The statistical theory of time-correlated measurements*. Review of Scientific Instruments, 1996. **67**(6): p. 2238-2246.
70. Charbon, E., *Single-photon imaging in complementary metal oxide semiconductor processes*. Philosophical Transactions of the Royal Society a-Mathematical Physical and Engineering Sciences, 2014. **372**(2012): p. 20130100.
71. Parmesan, L., N.A.W. Dutton, N.J. Calder, N. Krstajic, A.J. Holmes, L.A. Grant, and R.K. Henderson, *A 256×256 SPAD array with in-pixel Time to Amplitude Conversion for Fluorescence Lifetime Imaging Microscopy*, in *International Image Sensor Workshop*. 2015: Vaals, Netherlands.
72. Dutton, N.A.W., I. Gyongy, L. Parmesan, C.N. Gneccchi, B.R. Rae, S. Pellegrini, L.A. Grant, and R.K. Henderson, *A SPAD-based QVGA Image Sensor for Single Photon Counting and Quanta Imaging*. IEEE Transactions on Electron Devices, 2016. **63**(1): p. 189-196.
73. Pancheri, L. and D. Stoppa, *A SPAD-based Pixel Linear Array for High-Speed Time-Gated Fluorescence Lifetime Imaging*, in *2009 Proceedings of Esscirc*. 2009, IEEE: New York. p. 429-432.
74. Poland, S.P., N. Krstaji, J. Monypenny, S. Coelho, D. Tyndall, R.J. Walker, V. Devaues, J. Richardson, N. Dutton, P. Barber, D.D.-U. Li, K. Suhling, T. Ng, R.K. Henderson, and S.M. Ameer-Beg, *A high speed multifocal multiphoton fluorescence lifetime imaging microscope for live-cell FRET imaging*. Biomedical Optics Express, 2015. **6**(2): p. 277-296.
75. Poland, S.P., N. Krstajić, S. Coelho, D. Tyndall, R.J. Walker, V. Devaues, P.E. Morton, N.S. Nicholas, J. Richardson, D.D.-U. Li, K. Suhling, C.M. Wells, M. Parsons, R.K. Henderson, and S.M. Ameer-Beg, *Time-resolved multifocal multiphoton microscope for high speed FRET imaging in vivo*. Optics Letters, 2014. **39**(20): p. 6013-6016.
76. Esposito, A., *Beyond Range: Innovating Fluorescence Microscopy*. Remote Sensing, 2012. **4**(1): p. 111-119.
77. Charbon, E., M. Fishburn, R. Walker, R. Henderson, and C. Niclass, *SPAD-Based Sensors*, in *TOF Range-Imaging Cameras*, F. Remondino and D. Stoppa, Editors. 2013, Springer: Berlin, Heidelberg. p. 11-38.
78. Krstajić, N., S. Poland, J. Levitt, R. Walker, A. Erdogan, S. Ameer-Beg, and R.K. Henderson, *0.5 billion events per second time correlated single photon counting using CMOS SPAD arrays*. Optics Letters, 2015. **40**(18): p. 4305-4308.
79. Garrou, P., C. Bower, and P. Ramm, *Handbook of 3D Integration: Volume 1 - Technology and Applications of 3D Integrated Circuits*. 2011, Weinheim: John Wiley & Sons.
80. Bisogni, M.G. and M. Morrocchi, *Development of analog solid-state photo-detectors for Positron Emission Tomography*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2016. **809**: p. 140-148.
81. Pétusseau, A., P. Bruza, S. Tisa, S. Gioux, and B.W. Pogue, *SPAD based imaging of Cherenkov light in radiation therapy*. SPIE BiOS. Vol. 10862. 2019: SPIE.
82. Sekar, S.K.V., S. Mosca, S. Tannert, G. Valentini, F. Martelli, T. Binzoni, Y. Prokazov, E. Turbin, W. Zusratter, R. Erdmann, and A. Pifferi, *Time domain diffuse Raman spectrometer based on a TCSPC camera for the depth analysis of diffusive media*. Optics Letters, 2018. **43**(9): p. 2134-2137.



**Faculty of  
Natural & Mathematical  
Sciences**  
Department  
of Physics

**Professor Klaus Suhling**  
PhD (Strath) FlinstP CPhys  
Professor of Physics

Strand  
London WC2R 2LS  
UK  
Tel: +44 20 7848 2119  
Fax: +44 20 7848 2420  
Email: klaus.suhling@kcl.ac.uk



**University of London**

1<sup>st</sup> July 2019

Dear Dr. Demarteau,

Thank you very much for comments on our Ulitima conference proceedings manuscript.

*"I have read the paper and find that the paper provides a nice overview of the field. The paper shows the advantages and disadvantages of the various methods. Since this is a paper providing an overview of various methodologies, the fact that there are no original results presented is fine. All results are properly referenced. I suggest that the paper be published."*

Thank you very much for your kind comments.

*There are two small typos I found:*

*Line 140: fir -> for*

This has been fixed

*Line 144: "special cameras such as could allow" -> there is something missing in this sentence. Such as refers to examples that are not given. I suggest the author fix these two minor points and resubmit."*

Apologies – we have added the missing text, a reference to timepix and pimms cameras:

"...special cameras such as timepix [35, 36] or pimms [37] cameras, could allow the measurement of even faster sample decays."

References 35, 36 and 37 are already cited elsewhere in the text:

35. Fisher-Levine, M. and A. Nomerotski, *TimepixCam: a fast optical imager with time-stamping*. Journal of Instrumentation, 2016. **11**(03): p. C03016.

36. Zhao, A., M. van Beuzekom, B. Bouwens, D. Byelov, I. Chakaberia, C. Cheng, E. Maddox, A. Nomerotski, P. Svihra, J. Visser, V. Vrba, and T. Weinacht, *Coincidence velocity map imaging using Tpx3Cam, a time stamping optical camera with 1.5 ns timing resolution*. Review of Scientific Instruments, 2017. **88**(11): p. 113104.

37. John, J.J., M. Brouard, A. Clark, J. Crooks, E. Halford, L. Hill, J.W.L. Lee, A. Nomerotski, R. Pisarczyk, I. Sedgwick, C.S. Slater, R. Turchetta, C. Vallance, E. Wilman, B. Winter, and W.H. Yuen, *PImMS, a fast event-triggered monolithic pixel detector with storage of multiple timestamps*. Journal of Instrumentation, 2012. **7**(08): p. C08001.

We hope that the manuscript is now in a suitable form to be accepted for publication in NIMA, and we look forward to hearing from you.

Yours sincerely

A handwritten signature in blue ink, appearing to read "Klaus Suhling".

Klaus Suhling

1                   **Wide-field Time-correlated Single Photon Counting-based Fluorescence Lifetime**  
2                   **Imaging Microscopy**

3  
4                   Klaus Suhling<sup>\*a</sup>, Liisa M. Hirvonen<sup>a,h</sup>, Wolfgang Becker<sup>b</sup>, Stefan Smietana<sup>b</sup>, Holger Netz<sup>b</sup>, James Milnes<sup>c</sup>,  
5                   Thomas Conneely<sup>c</sup>, Alix Le Marois<sup>a,i</sup>, Ottmar Jagutzki<sup>d</sup>, Fred Festy<sup>e,j</sup>, Zdeněk Petrášek<sup>f</sup>, Andrew Beeby<sup>g</sup>

6  
7                   <sup>a</sup>Department of Physics, King’s College London, Strand, London WC2R 2LS, UK

8                   <sup>b</sup>Becker & Hickl GmbH, Nahmitzer Damm 30, 12277 Berlin, Germany

9                   <sup>c</sup>Photek Ltd, 26 Castleham Rd, St Leonards on Sea TN38 9NS, UK

10                   <sup>d</sup>Institut für Kernphysik, Max-von-Laue-Str. 1, 60438 Frankfurt, Germany

11                   <sup>e</sup>Biomaterials, Biomimetics and Biophotonics Research Group, Kings College London Dental Institute at Guys  
12                   Hospital, Kings Health Partners, Guys Dental Hospital, London Bridge, London SE1 9RT, UK

13                   <sup>f</sup>Institut für Biotechnologie und Bioprozesstechnik, Technische Universität Graz, Petersgasse 10-12/I, 8010  
14                   Graz, Austria

15                   <sup>g</sup>Department of Chemistry, University of Durham, Durham DH13LE, UK

16                   <sup>h</sup>present address: Randall Centre for Cell & Molecular Biophysics, New Hunt's House, King’s College London,  
17                   Guy’s Campus, London SE1 1UL, UK

18                   <sup>i</sup>present address: The Francis Crick Institute, 1 Midland Road, London NW1 1AT, UK

19                   <sup>j</sup>present address: Virtual AI, 24 Oakwood, Berkhamsted, Hertfordshire HP4 3NQ, UK  
20

21                   **Abstract**

22  
23                   Wide-field time-correlated single photon counting detection techniques, where the position and the arrival time of  
24                   the photons are recorded simultaneously using a camera, have made some advances recently. The technology and  
25                   instrumentation used for this approach is employed in areas such as nuclear science, mass spectroscopy and  
26                   positron emission tomography, but here, we discuss some of the wide-field TCSPC methods, for applications in  
27                   fluorescence microscopy. We describe work by us and others as presented in the Ulitima fast imaging and tracking  
28                   conference at the Argonne National Laboratory in September 2018, from phosphorescence lifetime imaging  
29                   (PLIM) microscopy on the microsecond time scale to fluorescence lifetime imaging (FLIM) on the nanosecond  
30                   time scale, and highlight some applications of these techniques.  
31

32                   **1 Introduction**

33  
34                   Optical microscopy is a widely-used tool for non-destructive and minimally invasive observation of living samples.  
35                   Fluorescence microscopy in particular allows the observation of cell dynamics and function in real time with  
36                   negligible cytotoxicity. It can also provide high sensitivity, down to the single molecule level, and high specificity.  
37                   In addition to localising fluorescent labels, the fluorescence can also be used for sensing the immediate  
38                   environment of the fluorophore, via its spectral properties, its polarization or fluorescence lifetime.[1] The  
39                   fluorescence lifetime is often used for this, as it is independent of the concentration of the fluorescence probe,  
40                   which is difficult to control in cells.  
41

42                   The fluorescence lifetime is the average time a fluorophore remains in the excited state, typically nanoseconds. It  
43                   can be a function of viscosity, temperature, pH, ion or glucose concentration, refractive index or polarity, and of  
44                   interaction with other molecules, e.g. due to Förster Resonance Energy Transfer (FRET), a widely used technique  
45                   to identify protein conformational changes or interactions. Phosphorescence, which originates from the probe’s  
46                   triplet state, is typically on the microsecond time scale, and is used to sense oxygen concentration, with  
47                   phosphorescence lifetime imaging (PLIM) being the imaging version of this approach. [2, 3] Fluorescence lifetime  
48                   imaging (FLIM) and PLIM can measure the lifetimes in every pixel, thus providing image contrast according to  
49                   the lifetime, which then provides contrast according to viscosity, oxygen or ion concentration or temperature,  
50                   depending on the type of fluorescent probe employed.  
51

52                   FLIM or PLIM is often carried out with time-correlated single photon counting (TCSPC) and confocal or  
53                   multiphoton beam scanning. TCSPC is effectively a delayed coincidence method,[4] the origin of which lies in  
54                   particle physics. It can be traced back to 1929, when Bothe and Kohlhörster used two Geiger counters separated  
55                   by spacers of varying thickness to study coincidences of penetrating charged particles in cosmic rays.[5] This was  
56                   followed by the first practical electronic coincidence circuit by Bruno Rossi in 1930, which became a precursor of  
57                   the AND logic gate in electronic circuits.[6] By the addition of a delay, this coincidence method evolved to measure  
58                   delayed coincidence, and thus provided the means for time-resolved measurements. A method to measure the  
59                   amplitude of the signal as a function of the delay between pulses by Rossi in 1942, a “time-circuit”, is now known  
60                   as the time-to-amplitude converter (TAC).[6] It became a popular method to measure short radioactive decays in

the 1950s,[7] and in 1961, Bollinger and Thomas generalised the scintillation measurements enabling TCSPC as we now know it, where the arrival time of a single photon is measured relative to an excitation pulse.[8] The accumulation of many single photons then represents the intensity decay of the sample, as long as no photons are lost due to pile-up, and the linearity between intensity and collected photons holds. The first reports that use TCSPC in the measurement of fluorescence decays appear in the early 1970s,[9-11] and TCSPC was soon widely used for time-resolved spectroscopy, and in particular the measurement of fluorescence lifetimes in solutions. Flashlamps used kHz repetition rates,[12] but lasers, with picosecond excitation pulses at MHz repetition rates, sped up the measurements significantly and advanced this field enormously.[13] In addition to scintillation and fluorescence measurements, TCSPC is also used for lidar [14] and optical tomography.[15]


TCSPC is a sensitive, precise, robust and mature technique to measure photon arrival times after an excitation pulse.[4] Its advantages stem from the digital nature of the technique, based on whether a single photon is detected, or not. It obeys well-defined Poisson statistics, which state that the experimental uncertainty is the square root of the number of counts, and that the signal-to-noise ratio increases with the measurement time, i.e. the number of counts. It also has a large dynamic range in time, from picoseconds to microseconds, and affords an easy visualisation of fluorescence decays. TCSPC has the highest signal-to-noise ratio of the standard time-resolved imaging methods,[16-18] and is precise enough to permit multi-exponential fluorescence lifetime fitting. By raster scanning the focal spot over the sample the image is created pixel by pixel, using single point detectors to perform TCSPC in each pixel.

While TCSPC is straight-forward to implement with scanning microscopy, there are a number of fluorescence microscopy methods that employ a camera, for example lightsheet microscopy, total internal reflection fluorescence (TIRF), supercritical angle fluorescence and super-resolution fluorescence microscopy methods based on localisation of fluorophores. To harness the advantages of TCSPC for these camera-based fluorescence microscopy methods, and to perform single photon sensitive wide-field FLIM or PLIM with these microscopy methods, a single photon sensitive camera with appropriate time resolution is required.[19-22] Sensitive detection is favoured, as fluorophores eventually bleach.[23] For microsecond time resolution PLIM, microchannel plate (MCP)-based photon counting image intensifiers with a phosphor screen and a fast camera can be used. For picosecond resolution FLIM, special read-out schemes for image intensifiers, e.g. charge division and propagation time techniques can be employed.

In addition to visible light, single photon detectors based on optoelectronic vacuum devices such as image intensifiers can also detect other types of electromagnetic radiation, including UV photons, x-rays and gamma rays, as well as particles, such as electrons, neutrons and ion fragments. MCPs are used in time-of-flight mass spectroscopy for detecting ion fragments, where molecules in vacuum chamber are ionised and broken up by short laser pulses.[24, 25] The fragments are accelerated by an electric field towards the MCP, and the position and arrival time contains information about the molecular fragments created. Furthermore, boron-doped MCPs are used for neutron detection and imaging,[26-28] as boron has a high neutron absorption cross section.[29] Timing of the neutron events allows distinction of cold and thermal neutrons, [27] and thus enables simultaneous imaging at different neutron energies. This yields different contrast, as absorption cross sections of the material in the object under study are a function of neutron energy. Moreover, in autoradiography, where radioactively labelled samples are imaged, MCPs can be used for sensitive detection of weak  $\beta$ -emitters such as  $^3\text{H}$ ,  $^{14}\text{C}$ , and  $^{35}\text{S}$ . [30] We note that for these approaches, once a particle has been converted into an electron in the MCP, the detection process is the same as for photons. Thus, development of wide-field detectors for FLIM and PLIM may also benefit these applications.

## 2. Wide-field time-correlated single photon counting imaging techniques and applications


Photon counting imaging, where the image is assembled from individual photons, effectively allows read-out noise free imaging. The approach goes back to the 1980s [31] and was, for example, used on the Faint Object Camera on the Hubble Space Telescope.[32] Note that the prime consideration at the time was sensitivity, not photon arrival timing. Modern MCP-based image intensifiers operated at saturated gain can perform photon counting imaging, and can either be equipped with a phosphor screen output, to be viewed with a camera, or with a position-sensitive anode.



## 2.1 Image intensifiers with a phosphor screen and camera readout for microsecond decay measurements

### 2.1.1 Photon arrival time obtained directly from the camera frame rate

Recent developments in imaging sensor technology have allowed complementary metal oxide semiconductor (CMOS) cameras to reach MHz frame rates.[33] These cameras can be used in combination with a photon counting image intensifier for TCSPC. [34] After each excitation pulse, a sequence of frames is acquired during the decay time of the probe, and this process is repeated until enough photons are collected so that a decay histogram is obtained for each pixel of the image, as shown in Figure 1. The time resolution of this approach is limited by the camera frame rate to the microsecond time scale. Although there is a trade-off between a high frame rate and the number of pixels that can be imaged, this technique enables the collection of up to hundreds of photons per frame, and even several photons after one excitation cycle per pixel as long as they arrive in different frames.[34] Latest developments of sensors that detect a signal above a certain threshold, e.g. timepix [35, 36] or pimms [37] cameras, can also be used for this purpose.[38] The details of the photon event intensity and area covered (number of camera pixels) is of no importance, only the fact where and when a photon event occurs, and therefore such an approach is feasible.




### 2.1.2 Photon arrival time obtained from imaging the phosphor decay

The phosphor decay of the intensifier output screen can be exploited to find the photon arrival time within the frame exposure time.[39] This effect, an afterglow, is usually undesired,[40] but can be put to good use. Matching the camera frame rate to the phosphor decay such that the photon events can be seen in several consecutive frames allows the photon arrival time within the camera exposure time to be found from the relative brightness of the photon event in successive frames, as illustrated in Figure 2. This approach is similar to dual exposure techniques for velocity map imaging in mass spectroscopy.[41, 42] The measurement of the photon arrival time from the phosphor decay can improve the time resolution beyond the inverse frame rate of the camera, and the lower frame rate increases the number of recorded pixels, thus allowing bigger field of view.[43, 44] Phosphorescent sample decays as short as 500 ns have been measured with a P20 phosphor and 300 kHz frame rate [45], as shown in Figure 3, but a combination of a faster phosphor and a faster frame rate, or special cameras such as timepix [35, 36] or pimms [37] cameras, could allow the measurement of even faster sample decays.

Camera-based wide-field TCSPC is especially well suited for sensitive measurements of phosphorescence lifetimes in the micro- and millisecond time region, i.e. PLIM.[2, 3] Camera-based methods enable the collection of hundreds of photons per excitation cycle, shortening the data acquisition time with long lifetime probes compared to single point scanning measurements. Lifetimes around 1  $\mu$ s have been measured with several transition metal probes using these techniques, including a ruthenium-based oxygen sensor in living cells, with total image acquisition times of just a few seconds.[33, 43] In fact, the best use of this approach is probably for



microsecond lifetime PLIM measurements due to the limited overall count rate of the image intensifier. It enables the collection of many photons per excitation pulse with low excitation power and without lengthy scanning.




## 2.2 Image intensifiers with an electronic anode readout for nanosecond decay measurements

For nanosecond fluorescence lifetime measurements, tens or hundreds of picosecond precision for photon arrival timing is essential. While the MCPs themselves are capable of timing the photon arrival with a precision of a few tens of picoseconds,[4] they are not capable of recording the photon's position without special read-outs. Different read-out architectures have been developed, including quadrant, wedge-and-strip, cross-strip and delay line anodes,[19, 20] where the position of the electron cloud is determined via a charge division approach, or via the propagation time along a delay line.[46] Although some of the read-out schemes can accommodate multiple photon events after one excitation cycle, i.e. the hexanode [47] or cross-strip read-outs,[19] the count rate is typically limited by the position readout electronics to a few 10s or 100s of kHz, rather than the need to avoid overlapping events. However, the advantage here is that sub-microwatt excitation powers are sufficient to generate photons, and allow continuous observation of living samples over days. Wide-field data collection allows the tracking of individual molecules or particle trajectories, e.g. single quantum dot tracking and FLIM has been demonstrated with a MCP and delay line anode detector.[48] Quadrant anodes [49, 50] have been applied to FLIM for the study of protein-protein interaction by FRET [51-53] and photosynthesis.[54, 55]

Wide-field TCSPC is especially useful for microscopy methods where the whole field of view is illuminated with a technique that provides depth discrimination. One of these techniques is TIRF microscopy, where the sample is excited by an evanescent wave only near (up to 100 nm) the coverslip. Total internal reflection (TIR)-FLIM has been demonstrated with quadrant anodes [56] and SPAD detectors.[57]

We have used a delay line anode detector (Photek) where the delay line is capacitively coupled to a resistive anode inside the tube and using an image charge technique,[20, 58] to perform TIR-FLIM.[59] Figure 4 shows wide-field TCSPC FLIM images of fixed HeLa cells acquired with this delay line anode detector, read out by conventional TCSPC timing boards (Becker & Hickl).[60] The sample was excited with a Horiba DeltaDiode picosecond laser (375 nm) at 10 MHz, and the photon count rate was around 80 kHz, with an acquisition time of 100s. The measured intensity shows the cell membrane stained with membrane dye laurdan only under TIRF illumination (Figure 4b), while the whole cell is visible under wide-field illumination (Figure 4c). The laurdan fluorescence lifetime is also shortened under TIRF illumination (Figure 4d) compared to wide-field illumination (Figure 4f); one contributory factor here is the proximity of the high refractive index glass coverslip which consequently shortens the fluorescence lifetime.[61] Figure 4g shows the measured instrument response function (IRF), and the measured time decays in a small area in Figure 4c,d. The IRF full width at half maximum (FWHM) is 344 ps.

The low excitation power used in wide-field TCSPC can help to minimise photodamage in living cells which is especially beneficial for observing dynamics in living cells over long periods.



### 2.3 Electron-bombarded sensors

Single photon detection is also possible with electron-bombarded (EB) sensors, where the photoelectrons from the photocathode are accelerated directly into a CCD or CMOS sensor.[62-66] Unlike MCPs where the statistical electron multiplication process creates a broad pulse height distribution, in EB sensors the photon event brightness depends on the gain voltage. Thus, by sweeping the gain voltage during the exposure time, it could be possible to obtain photon arrival time information from the photon event brightness, akin to a 2-dimensional streak camera.[62] The concept has been proposed and simulated, but not implemented; gain sweeping has not been possible with commercially available devices.

### 2.4 Single photon avalanche diode arrays

Another option to obtain a position-sensitive single photon detector is to build an array of single photon sensitive point detectors.[67-69] A single photon avalanche diode (SPAD, reverse biased above the diode breakdown voltage, Geiger mode) is a small all-solid state photon detector with a diameter of a few microns, and is capable of single photon detection with picosecond time resolution.[70] Single SPADs were first used for fast timing applications in the 1980s, and the implementation of SPADs in CMOS technology in 2003 enabled the development of SPAD arrays.[70] Unlike image intensifiers or EB sensors, SPADs do not require a high voltage or a vacuum, they are not damaged by high light levels, and they can be manufactured in arrays, 256×256 pixel[71] and 240×320 pixel SPAD arrays[72] have been reported.

SPAD arrays are a relatively new development in wide-field TCSPC. Initially, gated SPAD arrays for fluorescence lifetime were implemented [73], but now each photon can be timed individually, and in all pixels in parallel.[70] The big advantage of these developments in SPAD array detector technology is that it allows independent TCSPC in each pixel of a SPAD array simultaneously, e.g. in the 32 × 32 pixel megaframe chip, with a TDC in each pixel, with 55 ps resolution.[74, 75] They simultaneously deliver single photon sensitivity, tens of thousands of pixels spatial resolution and picosecond timing resolution.[70, 76, 77] The outstanding capability of enormous global count rates well into the gigahertz region,[78] which would allow the observation of fast cellular dynamics, is a big advantage of these devices.

The design of detectors and timing electronics on a single substrate inevitably provides compactness and large numbers of channels but compromises fill-factor and SPAD performance (jitter, photon detection efficiency, after pulsing and dark count). Nevertheless, SPAD array technology offers a huge advantage over existing FLIM detector technology. SPAD array detectors currently have a small fill factor (<10%), because the majority of the area of each pixel is occupied by electronic circuits to perform the timing, with only a small light-sensitive area dedicated to the detection of photons. Promising current developments in 3D stacking of integrated circuits [79] will ensure a fill factor >80%, a better time resolution and reduced jitter. Moreover, logic integration will scale up enormously enabling placement of field programmable gate arrays-like structures beneath the sensor. The development of 100% fill factor SPAD arrays will not only allow fast fluorescence lifetime measurements via wide-field TCSPC FLIM, the sensitivity and speed of this kind of detector could also benefit other applications, and this field continues to develop at a fast pace.[70, 77]. SPAD arrays are also used for positron emission tomography, where the picosecond timing capabilities can pinpoint the localization of the annihilation event more precisely than without timing, and thus increase the spatial resolution of the technique. [80] In addition, SPAD arrays have been used for detection of Cherenkov radiation in radiation therapy.[81] Range-finding is another application of SPAD arrays,[14] for example consumer electronics such as mobile phones have benefitted from TCSPC-based range finding, where sensitivity is not as big an issue as it is with FLIM.<sup>1</sup>

### 3. Conclusions

Fluorescence microscopy allows non-destructive and minimally invasive observation of living samples, and FLIM and PLIM allow the monitoring of the microenvironment of fluorescence and phosphorescence probes. A photon counting approach to FLIM and PLIM is particularly helpful, as it minimizes light exposure of the sample. It also is the best method to collect the fluorescence from the sample before the fluorophores are irreversibly bleached.[23] Wide-field TCSPC-based methods combine the advantages of single photon sensitivity and precision with wide-field data collection. This is important for implementation of specialised FLIM and PLIM microscopy methods that typically employ cameras, such as TIRF, lightsheet and others.

We have shown that wide-field time-correlated single photon counting based on an image intensifier with a phosphor screen and a fast CMOS camera can be employed for PLIM to map phosphorescent decays on a microsecond time scale. This can either be done with direct imaging of the photon events on the image intensifier's phosphor screen,[33] or by exploiting the invariant phosphor decay of the image intensifier screen for accurate timing of photon arrival well below the camera exposure time.[43] To image nanosecond fluorescence decays, we show that a crossed delay line anode detector read out with conventional TCSPC boards is feasible.[59] This approach retains all the advantages of TCSPC, and extends them to wide-field detection, serving essentially as a single photon sensitive camera with picosecond time resolution. Application of this approach to TIR FLIM using the membrane dye laurdan shows lifetime contrast between the plasma membrane and the interior membranes of the cell, as shown in figure 4. Photon counting approaches are particularly useful for fluorescence microscopy methods employing a camera to enable TCSPC-based FLIM, the FLIM method with the highest signal-to-noise ratio. The extremely low illumination intensity, distributed evenly over the field of view, is beneficial especially in life science applications where it allows long-term monitoring of living cells and organisms, while wide-field data collection enables the observation of cell dynamics and single particle tracking. A position-sensitive photon counting detector has recently been used for simultaneous acquisition of both spectral and temporal information of Raman photons from tissue phantoms,[82] showing the versatility of these types of detector.

Wide-field TCSPC methods are currently mainly based on MCP-based detectors, a mature technology used especially in astronomy and medical imaging. However, the development of SPAD arrays and their application to FLIM with the prospect of huge photon count rates continues at a rapid pace.

### 4. Acknowledgements

KS and LMH acknowledge BBSRC Sparking Impact and EPSRC Accelerated Impact funding for this work, as well as BBSRC grant BB/R004803.

<sup>1</sup> See [http://www.st.com/content/st\\_com/en/about/media-center/press-item.html/stmicroelectronics-proximity-sensor-solves-smartphone-hang-ups.html](http://www.st.com/content/st_com/en/about/media-center/press-item.html/stmicroelectronics-proximity-sensor-solves-smartphone-hang-ups.html) for information about proximity sensing in mobile phones using photon time-of-flight measurements with SPAD arrays.

## References

1. Suhling, K., L.M. Hirvonen, J.A. Levitt, P.-H. Chung, C. Tregidgo, A. Le Marois, D.A. Rusakov, K. Zheng, S. Ameer-Beg, S. Poland, S. Coelho, R. Henderson, and N. Krstajic, *Fluorescence lifetime imaging (FLIM): Basic concepts and some recent developments*. Medical Photonics, 2015. **27**: p. 3-40.
2. Shcheslavskiy, V.I., A. Neubauer, R. Bukowiecki, F. Dinter, and W. Becker, *Combined fluorescence and phosphorescence lifetime imaging*. Applied Physics Letters, 2016. **108**(9): p. 091111.
3. Baggaley, E., S.W. Botchway, J.W. Haycock, H. Morris, I.V. Sazanovich, J.A.G. Williams, and J.A. Weinstein, *Long-lived metal complexes open up microsecond lifetime imaging microscopy under multiphoton excitation: from FLIM to PLIM and beyond*. Chemical Science, 2014. **5**(3): p. 879-886.
4. O'Connor, D.V. and D. Phillips, *Time-correlated single-photon counting*. 1984, London: Academic Press.
5. Bothe, W. and W. Kolhörster, *Das Wesen der Höhenstrahlung*. Zeitschrift für Physik, 1929. **56**(11): p. 751-777.
6. Rossi, B., *Method of Registering Multiple Simultaneous Impulses of Several Geiger's Counters*. Nature, 1930. **125**: p. 636.
7. Bell, R.E., R.L. Graham, and H.E. Petch, *Design and use of a coincidence circuit of short resolving time*. Canadian Journal of Physics, 1952. **30**(1): p. 35-52.
8. Bollinger, L.M. and G.E. Thomas, *Measurement of the Time Dependence of Scintillation Intensity by a Delayed-Coincidence Method*. Review of Scientific Instruments, 1961. **32**(9): p. 1044-1050.
9. Binkert, T., H.P. Tschanz, and P.E. Zinsli, *The measurement of fluorescence decay curves with the single-photon counting method and the evaluation of rate parameters*. Journal of Luminescence, 1972. **5**(3): p. 187-217.
10. Lewis, C., W.R. Ware, L.J. Doemeny, and T.L. Nemzek, *The Measurement of Short-Lived Fluorescence Decay Using the Single Photon Counting Method*. Review of Scientific Instruments, 1973. **44**(2): p. 107-114.
11. Bachrach, R.Z., *A Photon Counting Apparatus for Kinetic and Spectral Measurements*. Review of Scientific Instruments, 1972. **43**(5): p. 734-737.
12. Birch, D.J.S. and R.E. Imhof, *Coaxial nanosecond flashlamp*. Review of Scientific Instruments, 1981. **52**(8): p. 1206-1212.
13. Phillips, D., *A lifetime in photochemistry; some ultrafast measurements on singlet states*. Proceedings of the Royal Society A: Mathematical, Physical and Engineering Sciences, 2016. **472**(2190): p. 20160102.
14. McCarthy, A., R.J. Collins, N.J. Krichel, V. Fernandez, A.M. Wallace, and G.S. Buller, *Long-range time-of-flight scanning sensor based on high-speed time-correlated single-photon counting*. Applied Optics, 2009. **48**(32): p. 6241-6251.
15. Schmidt, F.E.W., M.E. Fry, E.M.C. Hillman, J.C. Hebden, and D.T. Delpy, *A 32-channel time-resolved instrument for medical optical tomography*. Review of Scientific Instruments, 2000. **71**(1): p. 256-265.
16. Esposito, A., H.C. Gerritsen, and F.S. Wouters, *Optimizing frequency-domain fluorescence lifetime sensing for high-throughput applications: photon economy and acquisition speed*. Journal of the Optical Society of America A, 2007. **24**: p. 3261-3273.
17. Gratton, E., S. Breusegem, J. Sutin, Q. Ruan, and N. Barry, *Fluorescence lifetime imaging for the two-photon microscope: time-domain and frequency-domain methods*. Journal of Biomedical Optics, 2003. **8**(3): p. 381-390.
18. Philip, J. and K. Carlsson, *Theoretical investigation of the signal-to-noise ratio in fluorescence lifetime imaging*. Journal of the Optical Society of America A, 2003. **20**(2): p. 368-379.
19. Michalet, X., R.A. Colyer, G. Scalia, A. Ingargiola, R. Lin, J.E. Millaud, S. Weiss, O.H.W. Siegmund, A.S. Tremsin, J.V. Vallerger, A. Cheng, M. Levi, D. Aharoni, K. Arisaka, F. Villa, F. Guerrieri, F. Panzeri, I. Rech, A. Gulinatti, F. Zappa, M. Ghioni, and S. Cova, *Development of new photon-counting detectors for single-molecule fluorescence microscopy*. Philosophical Transactions of the Royal Society B-Biological Sciences, 2013. **368**(1611): p. 20120035.
20. Jagutzki, O., J.S. Lapington, L.B.C. Worth, U. Spillman, V. Mergel, and H. Schmidt-Böcking, *Position sensitive anodes for MCP read-out using induced charge measurement*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2002. **477**(1-3): p. 256-261.
21. Hartig, R., Y. Prokazov, E. Turbin, and W. Zusratter, *Wide-field fluorescence lifetime imaging with multi-anode detectors*, in *Fluorescence Spectroscopy and Microscopy: Methods and Protocols, Methods in Molecular Biology*, Y. Engelborghs and A.J.W.G. Visser, Editors. 2014, Springer. p. 457-480.
22. Prokazov, Y., E. Turbin, A. Weber, R. Hartig, and W. Zusratter, *Position sensitive detector for fluorescence lifetime imaging*. Journal of Instrumentation, 2014. **9**: p. C12015.

23. Zhao, Q.L., I.T. Young, and J.G.S. de Jong, *Photon budget analysis for fluorescence lifetime imaging microscopy*. Journal of Biomedical Optics, 2011. **16**(8): p. 086007.
24. Vallance, C., M. Brouard, A. Lauer, C.S. Slater, E. Halford, B. Winter, S.J. King, J.W.L. Lee, D.E. Pooley, I. Sedgwick, R. Turchetta, A. Nomerotski, J.J. John, and L. Hill, *Fast sensors for time-of-flight imaging applications*. Physical Chemistry Chemical Physics, 2014. **16**(2): p. 383-395.
25. Jungmann, J.H. and R.M.A. Heeren, *Detection systems for mass spectrometry imaging: A perspective on novel developments with a focus on active pixel detectors*. Rapid Communications in Mass Spectrometry, 2013. **27**(1): p. 1-23.
26. Tremsin, A.S., J.B. McPhate, J.V. Vallerga, O.H.W. Siegmund, W. Bruce Feller, E. Lehmann, A. Kaestner, P. Boillat, T. Panzner, and U. Filges, *Neutron radiography with sub-15  $\mu\text{m}$  resolution through event centroiding*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2012. **688**: p. 32-40.
27. Tremsin, A.S., J.B. McPhate, J.V. Vallerga, O.H.W. Siegmund, J.S. Hull, W.B. Feller, and E. Lehmann, *Detection efficiency, spatial and timing resolution of thermal and cold neutron counting MCP detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2009. **604**(1): p. 140-143.
28. Siegmund, O.H., J.V. Vallerga, A.S. Tremsin, J. McPhate, and B. Feller, *High spatial resolution neutron sensing microchannel plate detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2007. **576**(1): p. 178-182.
29. Fraser, G.W. and J.F. Pearson, *The direct detection of thermal neutrons by imaging microchannel-plate detectors*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 1990. **293**(3): p. 569-574.
30. Tarhoni, M.H., V. Vigneswara, M. Smith, S. Anderson, P. Wigmore, J.E. Lees, D.E. Ray, and W.G. Carter, *Detection, Quantification, and Microlocalisation of Targets of Pesticides Using Microchannel Plate Autoradiographic Imagers*. Molecules, 2011. **16**(10): p. 8535-8551.
31. Boksenberg, A., *Advances in Detectors for Astronomical Spectroscopy*. Philosophical Transactions of the Royal Society of London Series a-Mathematical Physical and Engineering Sciences, 1982. **307**(1500): p. 531-548.
32. Kröger, H.W., G.K. Schmidt, and N. Pailer, *Faint object camera: European contribution to the Hubble Space Telescope*. Acta Astronautica, 1992. **26**(11): p. 827-834.
33. Hirvonen, L.M., F. Festy, and K. Suhling, *Wide-field time-correlated single-photon counting (TCSPC) lifetime microscopy with microsecond time resolution*. Optics Letters, 2014. **39**(19): p. 5602-5605.
34. Sergeant, N., J.A. Levitt, M. Green, and K. Suhling, *Rapid wide-field photon counting imaging with microsecond time resolution*. Optics Express, 2010. **18**(24): p. 25292-25298.
35. Fisher-Levine, M. and A. Nomerotski, *TimepixCam: a fast optical imager with time-stamping*. Journal of Instrumentation, 2016. **11**(03): p. C03016.
36. Zhao, A., M. van Beuzekom, B. Bouwens, D. Byelov, I. Chakaberia, C. Cheng, E. Maddox, A. Nomerotski, P. Svihra, J. Visser, V. Vrba, and T. Weinacht, *Coincidence velocity map imaging using Tpx3Cam, a time stamping optical camera with 1.5 ns timing resolution*. Review of Scientific Instruments, 2017. **88**(11): p. 113104.
37. John, J.J., M. Brouard, A. Clark, J. Crooks, E. Halford, L. Hill, J.W.L. Lee, A. Nomerotski, R. Pisarczyk, I. Sedgwick, C.S. Slater, R. Turchetta, C. Vallance, E. Wilman, B. Winter, and W.H. Yuen, *PIImMS, a fast event-triggered monolithic pixel detector with storage of multiple timestamps*. Journal of Instrumentation, 2012. **7**(08): p. C08001.
38. Hirvonen, L.M., M. Fisher-Levine, K. Suhling, and A. Nomerotski, *Photon counting phosphorescence lifetime imaging with TimepixCam*. Review of Scientific Instruments, 2017. **88**(1): p. 013104.
39. Petrášek, Z. and K. Suhling, *Photon arrival timing with sub-camera exposure time resolution in wide-field time-resolved photon counting imaging*. Optics Express, 2010. **18**(24): p. 24888-24901.
40. Mainprize, J.G. and M.J. Yaffe, *The effect of phosphor persistence on image quality in digital x-ray scanning systems*. Medical Physics, 1998. **25**(12): p. 2440-2454.
41. Dinu, L., A.T.J.B. Eppink, F. Rosca-Pruna, H.L. Offerhaus, W.J. van der Zande, and M.J.J. Vrakking, *Application of a time-resolved event counting technique in velocity map imaging*. Review of Scientific Instruments, 2002. **73**(12): p. 4206-4213.
42. Strasser, D., X. Urbain, H.B. Pedersen, N. Altstein, O. Heber, R. Wester, K.G. Bhushan, and D. Zajfman, *An innovative approach to multiparticle three-dimensional imaging*. Review of Scientific Instruments, 2000. **71**(8): p. 3092-3098.
43. Hirvonen, L.M., Z. Petrášek, A. Beeby, and K. Suhling, *Sub- $\mu\text{s}$  time resolution in wide-field time-correlated single photon counting microscopy obtained from the photon event phosphor decay*. New Journal of Physics, 2015. **17**(2): p. 023032.



44. Hirvonen, L.M., Z. Petrášek, and K. Suhling, *Wide-field time-correlated single photon counting (TCSPC) microscopy with time resolution below the frame exposure time*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2015. **787**: p. 1-5.
45. Hirvonen, L.M., Z. Petrášek, A. Beeby, and K. Suhling, *Microsecond wide-field TCSPC microscopy based on an ultra-fast CMOS camera*. 2015, SPIE Proc 9329. p. 932939.
46. Hirvonen, L.M. and K. Suhling, *Wide-field TCSPC: methods and applications*. Measurement Science and Technology, 2017. **28**(1): p. 012003.
47. Jagutzki, O., A. Cerezo, A. Czasch, R. Dörner, M. Hattass, M. Huang, V. Mergel, U. Spillmann, K. Ullmann-Pfleger, T. Weber, H. Schmidt-Böcking, and G.D.W. Smith, *Multiple hit readout of a microchannel plate detector with a three-layer delay-line anode*. IEEE Transactions on Nuclear Science, 2002. **49**(5): p. 2477-2483.
48. Michalet, X., R.A. Colyer, J. Antelman, O.H.W. Siegmund, A. Tremsin, J.V. Vallerga, and S. Weiss, *Single-Quantum Dot Imaging with a Photon Counting Camera*. Current Pharmaceutical Biotechnology, 2009. **10**: p. 543-558.
49. Kemnitz, K., L. Pfeifer, and M.R. Ainbund, *Detector for multichannel spectroscopy and fluorescence lifetime imaging on the picosecond timescale*. Nuclear Instruments and Methods in Physics Research Section A, 1997. **387**(1/2): p. 86-87.
50. Emiliani, V., D. Sanvitto, M. Tramier, T. Piolot, Z. Petrášek, K. Kemnitz, C. Duneux, and M. Coppey-Moisán, *Low-intensity two-dimensional imaging of fluorescence lifetimes in living cells*. Applied Physics Letters, 2003. **83**(12): p. 2471-2473.
51. Vitali, M., F. Picazo, Y. Prokazov, A. Duci, E. Turbin, C. Gotze, J. Llopis, R. Hartig, A.J.W.G. Visser, and W. Zschurrer, *Wide-Field Multi-Parameter FLIM: Long-Term Minimal Invasive Observation of Proteins in Living Cells*. Plos One, 2011. **6**(2): p. e15820.
52. Ramadass, R., D. Becker, M. Jendrach, and J. Bereiter-Hahn, *Spectrally and spatially resolved fluorescence lifetime imaging in living cells: TRPV4-microfilament interactions*. Archives of Biochemistry and Biophysics, 2007. **463**(1): p. 27-36.
53. Gautier, I., M. Tramier, C. Durieux, J. Coppey, R.B. Pansu, J.C. Nicolas, K. Kemnitz, and M. Coppey-Moisán, *Homo-FRET microscopy in living cells to measure monomer-dimer transition of GFP-tagged proteins*. Biophysical Journal, 2001. **80**(6): p. 3000-3008.
54. Petrášek, Z., H.J. Eckert, and K. Kemnitz, *Wide-field photon counting fluorescence lifetime imaging microscopy: application to photosynthesizing systems*. Photosynthesis Research, 2009. **102**(2-3): p. 157-168.
55. Petrášek, Z., F.J. Schmitt, C. Theiss, J. Huyer, M. Chen, A. Larkum, H.J. Eichler, K. Kemnitz, and H.J. Eckert, *Excitation energy transfer from phycobiliprotein to chlorophyll d in intact cells of Acaryochloris marina studied by time-and wavelength-resolved fluorescence spectroscopy*. Photochemical & Photobiological Sciences, 2005. **4**(12): p. 1016-1022.
56. Giraud, G., H. Schulze, T. Bachmann, C. Campbell, A. Mount, P. Ghazal, M. Khondoker, A. Ross, S. Ember, I. Ciani, C. Tlili, A. Walton, J. Terry, and J. Crain, *Fluorescence Lifetime Imaging of Quantum Dot Labeled DNA Microarrays*. International Journal of Molecular Sciences, 2009. **10**(4): p. 1930-1941.
57. Giraud, G., H. Schulze, D.U. Li, T.T. Bachmann, J. Crain, D. Tyndall, J. Richardson, R. Walker, D. Stoppa, E. Charbon, R. Henderson, and J. Arlt, *Fluorescence lifetime biosensing with DNA microarrays and a CMOS-SPAD imager*. Biomedical Optics Express, 2010. **1**(5): p. 1302-1308.
58. Milnes, J., J.S. Lapington, O. Jagutzki, and J. Howorth, *Image charge multi-role and function detectors*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2009. **604**(1-2): p. 218-220.
59. Hirvonen, L.M., W. Becker, J. Milnes, T. Conneely, S. Smietana, A. Le Marois, O. Jagutzki, and K. Suhling, *Picosecond wide-field time-correlated single photon counting fluorescence microscopy with a delay line anode detector*. Applied Physics Letters, 2016. **109**: p. 071101.
60. Becker, W., L.M. Hirvonen, J. Milnes, T. Conneely, O. Jagutzki, H. Netz, S. Smietana, and K. Suhling, *A wide-field TCSPC FLIM system based on an MCP PMT with a delay-line anode*. Review of Scientific Instruments, 2016. **87**: p. 093710.
61. Tregidgo, C., J.A. Levitt, and K. Suhling, *Effect of refractive index on the fluorescence lifetime of green fluorescent protein*. Journal of Biomedical Optics, 2008. **13**: p. 031218.
62. Hirvonen, L.M., S. Jiggins, N. Sergeant, G. Zanda, and K. Suhling, *Photon counting imaging with an electron-bombarded CCD: towards a parallel-processing photoelectronic time-to-amplitude converter*. Rev Sci Instrum, 2014. **85**(12): p. 123102.
63. Hirvonen, L.M., S. Jiggins, N. Sergeant, G. Zanda, and K. Suhling, *Photon counting imaging with an electron-bombarded CCD: Towards wide-field time-correlated single photon counting (TCSPC)*.

- Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2015. **787**: p. 323-327.
64. Cajgfinger, T., A. Dominjon, and R. Barbier, *Single photon detection and localization accuracy with an ebCMOS camera*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 2015. **787**: p. 176-181.
65. Buontempo, S., et al., *The Megapixel EBCCD: A high-resolution imaging tube sensitive to single photons*. Nuclear Instruments & Methods in Physics Research Section A: Accelerators Spectrometers Detectors and Associated Equipment, 1998. **413**(2-3): p. 255-262.
66. Hirvonen, L.M. and K. Suhling, *Photon Counting Imaging with an Electron-Bombarded Pixel Image Sensor*. Sensors, 2016. **16**(5): p. 617.
67. Rinnenthal, J.L., C. Börnchen, H. Radbruch, V. Andresen, A. Mossakowski, V. Siffrin, T. Seelemann, H. Spiecker, I. Moll, J. Herz, A.E. Hauser, F. Zipp, M.J. Behne, and R. Niesner, *Parallelized TCSPC for Dynamic Intravital Fluorescence Lifetime Imaging: Quantifying Neuronal Dysfunction in Neuroinflammation*. Plos One, 2013. **8**(4): p. e60100.
68. McLoskey, D., D.J.S. Birch, A. Sanderson, K. Suhling, E. Welch, and P.J. Hicks, *Multiplexed single-photon counting. I. A time-correlated fluorescence lifetime camera*. Review of Scientific Instruments, 1996. **67**(6): p. 2228-2237.
69. Suhling, K., D. McLoskey, and D.J.S. Birch, *Multiplexed single-photon counting. II. The statistical theory of time-correlated measurements*. Review of Scientific Instruments, 1996. **67**(6): p. 2238-2246.
70. Charbon, E., *Single-photon imaging in complementary metal oxide semiconductor processes*. Philosophical Transactions of the Royal Society a-Mathematical Physical and Engineering Sciences, 2014. **372**(2012): p. 20130100.
71. Parmesan, L., N.A.W. Dutton, N.J. Calder, N. Krstajic, A.J. Holmes, L.A. Grant, and R.K. Henderson, *A 256×256 SPAD array with in-pixel Time to Amplitude Conversion for Fluorescence Lifetime Imaging Microscopy*, in *International Image Sensor Workshop*. 2015: Vaals, Netherlands.
72. Dutton, N.A.W., I. Gyongy, L. Parmesan, C.N. Gneccchi, B.R. Rae, S. Pellegrini, L.A. Grant, and R.K. Henderson, *A SPAD-based QVGA Image Sensor for Single Photon Counting and Quanta Imaging*. IEEE Transactions on Electron Devices, 2016. **63**(1): p. 189-196.
73. Pancheri, L. and D. Stoppa, *A SPAD-based Pixel Linear Array for High-Speed Time-Gated Fluorescence Lifetime Imaging*, in *2009 Proceedings of Esscirc*. 2009, IEEE: New York. p. 429-432.
74. Poland, S.P., N. Krstaji, J. Monypenny, S. Coelho, D. Tyndall, R.J. Walker, V. Devaues, J. Richardson, N. Dutton, P. Barber, D.D.-U. Li, K. Suhling, T. Ng, R.K. Henderson, and S.M. Ameer-Beg, *A high speed multifocal multiphoton fluorescence lifetime imaging microscope for live-cell FRET imaging*. Biomedical Optics Express, 2015. **6**(2): p. 277-296.
75. Poland, S.P., N. Krstajić, S. Coelho, D. Tyndall, R.J. Walker, V. Devaues, P.E. Morton, N.S. Nicholas, J. Richardson, D.D.-U. Li, K. Suhling, C.M. Wells, M. Parsons, R.K. Henderson, and S.M. Ameer-Beg, *Time-resolved multifocal multiphoton microscope for high speed FRET imaging in vivo*. Optics Letters, 2014. **39**(20): p. 6013-6016.
76. Esposito, A., *Beyond Range: Innovating Fluorescence Microscopy*. Remote Sensing, 2012. **4**(1): p. 111-119.
77. Charbon, E., M. Fishburn, R. Walker, R. Henderson, and C. Niclass, *SPAD-Based Sensors*, in *TOF Range-Imaging Cameras*, F. Remondino and D. Stoppa, Editors. 2013, Springer: Berlin, Heidelberg. p. 11-38.
78. Krstajić, N., S. Poland, J. Levitt, R. Walker, A. Erdogan, S. Ameer-Beg, and R.K. Henderson, *0.5 billion events per second time correlated single photon counting using CMOS SPAD arrays*. Optics Letters, 2015. **40**(18): p. 4305-4308.
79. Garrou, P., C. Bower, and P. Ramm, *Handbook of 3D Integration: Volume 1 - Technology and Applications of 3D Integrated Circuits*. 2011, Weinheim: John Wiley & Sons.
80. Bisogni, M.G. and M. Morrocchi, *Development of analog solid-state photo-detectors for Positron Emission Tomography*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2016. **809**: p. 140-148.
81. Pétusseau, A., P. Bruza, S. Tisa, S. Gioux, and B.W. Pogue, *SPAD based imaging of Cherenkov light in radiation therapy*. SPIE BiOS. Vol. 10862. 2019: SPIE.
82. Sekar, S.K.V., S. Mosca, S. Tannert, G. Valentini, F. Martelli, T. Binzoni, Y. Prokazov, E. Turbin, W. Zusratter, R. Erdmann, and A. Pifferi, *Time domain diffuse Raman spectrometer based on a TCSPC camera for the depth analysis of diffusive media*. Optics Letters, 2018. **43**(9): p. 2134-2137.